

# Certificate of Quality

**Product Name:** Acetate Differential Agar (slant) 20/PK  
**Lot Number:** 694110

**Product Number:** R060022  
**Expiration Date:** 2009-08-18  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Escherichia coli	9339	Growth With Color Change to Blue
Shigella sonnei	9290	No Growth

**Physical Characteristics:**  
Appearance: Green, Clear Agar Slant  
pH @ 20-25°C: 6.7 +/- 0.2

## **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor