

# Certificate of Quality

**Product Name:** Anaer Medium I-15/PK  
**Lot Number:** 701964

**Product Number:** R01099  
**Expiration Date:** 2008-11-20  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Bacteroides fragilis	25285	Good Growth
Fusobacterium nucleatum	25586	Growth
Peptostreptococcus anaerobius	27337	Good Growth
Clostridium perfringens	13124	Good Growth
Escherichia coli	25922	Good Growth
Staphylococcus aureus	25923	Good Growth

#### **Physical Characteristics:**

Appearance: Red, Opaque Agar  
pH @ 20-25°C: 7.4 +/- 0.2

#### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor