

Certificate of Quality

Product Name: Tergitol 7 Agar w/o TTC 10/PK
Lot Number: 706680

Product Number: R01877
Expiration Date: 2008-12-09
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Proteus mirabilis	12453	Blue to Green Colonies, No Swarming
Proteus vulgaris	6380	Blue to Green Colonies, No Swarming
Salmonella typhimurium	14028	Blue to Green Colonies
Escherichia coli	25922	Yellow Colonies
Staphylococcus aureus	25923	Marked to Complete Inhibition

Physical Characteristics:
Appearance: Light Green, Clear Agar
pH @ 20-25°C: 6.9 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor