

Certificate of Quality

Product Name: Lysine Iron Agar (LIA) (slant) 20/PK
Lot Number: 712553

Product Number: R061302
Expiration Date: 2009-07-16
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Shigella sonnei	9290	Purple Slant, Yellow Butt, H2S (-)
Escherichia coli	25922	Purple Slant, Purple Butt, H2S (-)
Proteus mirabilis	12453	Red Slant, Yellow Butt, H2S (-)
Salmonella typhimurium	14028	Purple Slant, Purple Butt, H2S (+) Black Production

Physical Characteristics:
Appearance: Purple, Clear Agar Slant
pH @ 20-25°C: 6.7 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor