



## Certificate of Quality

**Product Name:** Fluid Thio Med (9ml) 100/PK  
**Lot Number:** 728930

**Product Number:** R064690  
**Expiration Date:** 2009-09-08  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge       | ATCC  | Performance Criteria |
|--------------------------|-------|----------------------|
| * Staphylococcus aureus  | 6538  | Growth               |
| * Clostridium sporogenes | 11437 | Growth               |
| * Pseudomonas aeruginosa | 9027  | Growth               |
| ** Staphylococcus aureus | 25923 | Good Growth          |
| ** Bacteroides fragilis  | 25285 | Good Growth          |

- \* Product Tested According To USP Growth Promotion Testing Procedure
- \*\* Product Tested According To NCCLS Growth Promotion Testing Procedure

### Physical Characteristics:

Appearance: Light to Medium Amber, Clear Broth With Visible Agar  
pH @ 20-25°C: 7.1 +/- 0.2

### Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor