



Certificate of Analysis

Human IL-10 ELISA Kit

Product No. KHC0101C
Lot No. 73697177A
Expiration Date 31May2015

Part Number	Lot Number	Description	Part Number	Lot Number	Description
SD019	7B5/1	Standard	PP019	13G8/1	Ab Coated Plates
DS83	13I32/1	Standard Diluent	WB01	As Supplied	Wash Buffer
BC019	14B30/1	Biotin Conjugate	SB01	13K48/2	Chromogen
SV019	14B7/1	Streptavidin-HRP	SS01	As Supplied	Stop Solution
DV82	13J8/1	HRP Diluent			

Parameter	Standard Conc. pg/mL	Avg. O.D.
Standard 1	500	3196.0
Standard 2	250	1942.0
Standard 3	125	1189.0
Standard 4	62.5	629.0
Standard 5	31.2	359.3
Standard 6	15.6	219.7
Standard 7	7.8	136.7
Standard 8	0.0	73.7

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com

Michael Grossman
Sr. Manager, GOS – Quality
Issued on 15 Apr 2014