

Certificate of Quality

Product Name: MH Agar w/5% Sheep Blood (100mm) 10/PK
Lot Number: 743039

Product Number: R01622
Expiration Date: 2009-04-07
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Streptococcus pneumoniae	49619	All Zones within NCCLS Guidelines

This product has been tested and is compliant with the performance criteria in the NCCLS document, performance standards for antimicrobial disk susceptibility tests.

Physical Characteristics:

Appearance:	Red, Opaque Agar
pH @ 20-25°C:	7.3 +/- 0.1

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor