



Certificate of Quality

Product Name: NLA - Bordet Gengou Agar (150mm) 10/PK
Lot Number: 751308

Product Number: R110810
Expiration Date: 2009-04-16
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|-----------------------|-------|----------------------|
| Bordetella pertussis | 12742 | Good Growth |
| Streptococcus sanguis | 10556 | Good Growth |

Physical Characteristics:

| | |
|---------------|------------------|
| Appearance: | Red, Opaque Agar |
| pH @ 20-25°C: | 6.7 +/- 0.2 |

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor