

Certificate of Quality

Product Name: Cystine Tellurite Blood Agar 10/PK
Lot Number: 753163

Product Number: R01348
Expiration Date: 2009-04-23
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Corynebacterium diphtheriae	13812	Good Growth, Black Colonies
Staphylococcus aureus	25923	No Growth
Streptococcus sanguis	10556	No Growth

Physical Characteristics:

Appearance: Red, Opaque Agar w/ Precipitate
pH @ 20-25°C: 7.4 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor