



Certificate of Quality

Product Name: ^RO^ M4RT 100ML
Lot Number: 755179

Product Number: R12625
Expiration Date: 2010-09-02
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge **ATCC** **Performance Criteria**

Maintenance:

Chlamydia trachomatis	VR-880	Pass
Herpes Simplex type 2	VR-734	Pass
Cytomegalovirus	VR-977	Pass
Resp. Syncytial Virus	VR-1302	Pass

Toxicity:

MRC5 Cells	CCL-171	Pass
VERO Cells	CCL-81	Pass
McCoy cells	RLL-169	Pass

Physical Characteristics:

Appearance:	Orange-Pink Clear Broth
pH @ 20-25°C:	7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor