

Certificate of Quality

Product Name: Dermatophyte Test Medium (26ml) 10/PK
Lot Number: 757666

Product Number: R01365
Expiration Date: 2009-06-04
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Aspergillus niger	16404	Inhibited
Candida albicans	10231	Good Growth
Trichophyton mentagrophytes	9533	Good Growth, Red Zone
Cryptococcus neoformans	34877	No Growth
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth
Pseudomonas aeruginosa	27853	No Growth

Physical Characteristics:

Appearance: Bright Orange, Clear Agar
pH @ 20-25°C: 5.5 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor