

# Certificate of Quality

**Product Name:** BHI Agar w/SB,Chlor,Gent (26ml) 10/PK  
**Lot Number:** 769551

**Product Number:** R01144  
**Expiration Date:** 2009-06-30  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Candida albicans	10231	Good Growth
Cryptococcus neoformans	14116	Good Growth
Trichophyton mentagrophytes	9533	Good Growth
Escherichia coli	25922	No Growth
Staphylococcus aureus	25923	No Growth
Pseudomonas aeruginosa	27853	No Growth

#### **Physical Characteristics:**

Appearance: Red, Opaque Agar  
pH @ 20-25°C: 7.4 +/- 0.2

#### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor