

Certificate of Quality

Product Name: TSA 10/PK
Lot Number: 772473

Product Number: R01920
Expiration Date: 2009-07-22
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	8739	Good Growth
Staphylococcus aureus	6538	Good Growth
Pseudomonas aeruginosa	9027	Good Growth
Bacillus subtilis	6633	Good Growth
Candida albicans	10231	Good Growth
Aspergillus niger	16404	Good Growth

Physical Characteristics:

Appearance: Light to Medium Amber, Clear Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor