

Certificate of Quality

Product Name: Anaer Reducible CNA Bld Agar 15/PK
Lot Number: 788957

Product Number: R01065
Expiration Date: 2009-08-31
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|-------------------------------|-------------|-----------------------------|
| Bacteroides fragilis | 25285 | Good Growth |
| Peptostreptococcus anaerobius | 27337 | Good Growth |
| Staphylococcus aureus | 25923 | Good Growth |
| Escherichia coli | 25922 | Inhibited |

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor