

Certificate of Quality

Product Name: C. albicans ATCC 10231 Q+ KT/100 TST
Lot Number: 796763

Product Number: Q1503C
Expiration Date: 2011-01-31
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Purity:

Standardized aliquots of the rehydrated product are inoculated onto nonselective media and examined for pure growth following the appropriate incubation. Selective and Differential media are also tested where applicable.

Viability And Quantification:

Each organism is recovered from the preserved state within the required time frame and at an acceptable level. Passage number is stated as the current preserved state.

Macroscopic And Microscopic Morphology:

Colony morphology is consistent with documented referenced description. Traditional staining is performed.

Biochemical Analysis:

Organism exhibits characteristic biochemical and/or enzymatic reactions. Automated and/or conventional testing was performed and results were within established limits. Antimicrobial testing performed where applicable. Results within expected ranges.

CFU/ 0.1ml: <100
Gram Reaction: Gram Positive Yeast


Passage: 2
Biochemical Profile: Vitek YBC

Appearance: Preserved Gel Matrix suspended in vial cap
pH: N/A

Colony Count: Not less than 5 samples each were plated to the following:
Trypticase Soy Agar, spread plates, incubated at 35°C for 24 - 48 hours.

Colony forming units as noted above are specified under defined growing conditions. Results may vary with more inhibitory or selective media or with the same medium of inferior quality.

Signed



Product Performance Technologist