



Certificate of Quality

Product Name: MacConkey CT-Sorbitol 10/PK
Lot Number: 810016

Product Number: R110241
Expiration Date: 2009-11-16
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	43888	Colorless Colonies
Escherichia coli	25922	Complete Inhibition to Growth, Pink Colonies
Proteus mirabilis	12453	Inhibited

Physical Characteristics:

Appearance: Clear Agar, Pink
pH @ 20-25°C: 7.1 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor