

Certificate of Quality

Product Name: Blood Agar,5% Sheep Blood 100/PK
Lot Number: 814639

Product Number: R01202
Expiration Date: 2009-11-17
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Streptococcus pyogenes	19615	Good Growth, Beta Hemolysis, Camp Test Negative
Streptococcus agalactiae	12386	Good Growth, Beta Hemolysis, Camp Test Positive
Enterococcus faecalis	29212	Good Growth, Gamma Hemolysis
Streptococcus pneumoniae	6305	Good Growth, Alpha Hemolysis
Listeria monocytogenes	7646	Good Growth, Beta Hemolysis
Escherichia coli	25922	Good Growth
Staphylococcus aureus	25923	Good Growth
Moraxella catarrhalis	25238	Good Growth

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor