

Certificate of Quality

Product Name: EMB Agar,Levine 10/PK
Lot Number: 818262

Product Number: R01400
Expiration Date: 2009-12-02
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	25922	Blue-Black Colonies, Green Metallic Sheen
Proteus mirabilis	12453	Clear Colonies, No Swarming
Salmonella typhimurium	14028	Clear Colonies
Staphylococcus aureus	25923	Marked Inhibition
Enterococcus faecalis	29212	Inhibited

Physical Characteristics:
Appearance: Brown-Purple, Clear Agar
pH @ 20-25°C: 7.1 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor