



## Certificate of Quality

**Product Name:** Sab Dex Agar, Emmons (slant) 20/PK  
**Lot Number:** 819835

**Product Number:** R064422  
**Expiration Date:** 2010-09-22  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Candida albicans	10231	Good Growth
Cryptococcus neoformans	34877	Good Growth
Trichophyton mentagrophytes	9533	Good Growth

### **Physical Characteristics:**

Appearance: Light to Medium Amber, Clear Agar Slant  
pH @ 20-25°C: 7.0 +/- 0.2

### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor