

## Certificate of Analysis

**Product Name:** Streptex Acid Extraction KT/50TST  
**Product Code:** R30951301  
**Lot Number:** 821924  
**Expiration Date:** 31.12.2011  
**Date of Manufacture:** 07.01.2010

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74AZL59A	Extraction Reagent 1	822707	31.12.2011
W74AZL59B	Extraction Rgt 2	822752	31.12.2011
W74AZL59C	Extraction Reagent 3	822754	31.12.2011

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Efficacy of Extraction Reagents 1, 2 and 3	Each Latex must show obvious agglutination with its homologous extracts	Conforms
	Each Latex must not show agglutination with the heterologous extracts, nor with the extraction fluid from the uninoculated tube although trace granularity is acceptable	Conforms



PAUL BROWNING  
QA MANAGER