

Certificate of Quality

Product Name: Anaer (CDC) Blood Agar 15/PK
Lot Number: 840913

Product Number: R01039
Expiration Date: 2010-01-13
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bacteroides fragilis	25285	Good Growth
Clostridium perfringens	13124	Good Growth
Fusobacterium nucleatum	25586	Good Growth
Peptostreptococcus anaerobius	27337	Good Growth
Escherichia coli	25922	Good Growth, Beta Hemolysis
Staphylococcus aureus	25923	Good Growth

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.5 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor