

Certificate of Quality

Product Name: Anaer Reducible Blood Agar 15/PK
Lot Number: 850436

Product Number: R01061
Expiration Date: 2010-02-23
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|-------------------------------|-------------|-----------------------------|
| Bacteroides fragilis | 25285 | Good Growth |
| Clostridium perfringens | 13124 | Good Growth, Beta Hemolysis |
| Fusobacterium nucleatum | 25586 | Good Growth |
| Peptostreptococcus anaerobius | 27337 | Good Growth |
| Prevotella melaninogenica | 25845 | Good Growth |
| Escherichia coli | 25922 | Good Growth |
| Staphylococcus aureus | 25923 | Good Growth |

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor