



## CERTIFICATE OF ANALYSIS

Page 1 of 1

**PRODUCT** TV5016D  
MAXIMUM RECOVERY DILUENT 9ML

**LOT NUMBER** 874279

**EXPIRY DATE** 2011.02.25

### TYPICAL FORMULATION

	grams per litre
Peptone	1.00000
Sodium chloride	8.50000

General Characteristics	Results	Specification
Colour	Conforms	Colourless
Appearance	Conforms	Transparent
pH	7.1	6.8 - 7.2
Packaging/Presentation	Conforms	Conforms to specification
Fill Volume (ML)	Conforms	8.82 - 9.18
Sterility @ 25 & 36 ± 1°C for 72 hours	Conforms	No Contamination

Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®25923	55	55	cfu +/- 50% of original count
Escherichia coli ATCC®25922	36	51	cfu +/- 50% of original count

Tested in accordance with CEN ISO/TS 11133-2.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature