



Certificate of Quality

Product Name: EC Medium w/MUG (10ml) 20/PK
Lot Number: 880979

Product Number: R07655
Expiration Date: 2011-03-15
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Escherichia coli	25922	Good Growth, Blue Fluorescence
Enterococcus faecalis	29212	No Growth

Physical Characteristics:

Appearance: Light to Medium Amber, Clear Broth
pH @ 20-25°C: 6.9 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor