



## Certificate of Quality

**Product Name:** ^MK^ 10/PK SP4 GLU AGAR 60MM  
**Lot Number:** 930444

**Product Number:** R20279  
**Expiration Date:** 2010-10-21  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Challenge Organisms	ATCC	Performance Criteria
*Mycoplasma pneumoniae	15531	Good Growth
**Escherichia coli	25922	No Growth
**Staphylococcus aureus	25923	No Growth
Mycoplasma arginini	23838	Good Growth

\* Product tested according to USP growth promotion testing procedure \*

\*\* Product tested according to CLSI growth promotion testing procedure \*\*

### Microbial Load

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

### Physical Characteristics:

Appearance: Orange-red, Clear Agar

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor