

## **CERTIFICATE OF ANALYSIS**

PRODUCT NAME: Precise Tris-Glycine Gels, 10cm x 8cm, 10%, 15-Well

**PRODUCT NUMBER: 25271** 

LOT NUMBER: QE0024311

## SPECIFICATIONS:

RETARDANT FACTOR:

(Rf): value for Ovostatin: 0.02 - 0.09 (Rf): value for Lysozyme: 0.86 - 0.98

(Rf): difference between Ovotransferrin & Ovalbumin: ≥ 0.18

**INSPECTION: Pass** 

**RESULTS:** 

RETARDANT FACTOR:

(Rf): value for Ovostatin: 0.05 (Rf): value for Lysozsyme: 0.96

(Rf): difference between Ovotransferrin & Ovalbumin: 0.21

**INSPECTION: Pass** 

Publish Date: 4/23/2015

Liam F. Garrity, Senior Manager, Analytical Services

Mint. Cam

Laboratory

[Rev.001]