



CERTIFICATE OF ANALYSIS

PRODUCT MT1907
SABOURAUD DEXTROSE AGAR SLANTS PCK 25
15 ML 28 X 95MM PCK 25

LOT NUMBER 0263

EXPIRY DATE 22.03.2011
(dd.mm.yyyy)

Delivery/Customer information

Date Printed (dd.mm.yyyy)

11.11.2013

Delivery No.

Customer

Customer Order number

Physical Characteristics	Results	Specification
Appearance	Amber slant	Conforms to Specification
pH	5.40	5.40 - 5.80
Packaging and Presentation	Conforms	Conforms to specification
Sterility @ 35 °C ± 2°C for 3 days	Conforms	Conforms to specification

Microbiological Performance		
Organism Tested	Expected Results	Specification
A. brasiliensis frm. niger ATCC® 16404	Good to Excellent growth	Conforms to Specification
Trichophyton mentagrophytes ATCC® 9533	Good to Excellent growth	Conforms to Specification
Trichophyton rubrum ATCC® 28188	Good to Excellent growth	Conforms to Specification
Candida albicans ATCC® 10231	Good to Excellent growth	Conforms to Specification

The Quality Assurance of this medium has been performed in accordance with CLSI document M22-A3. All quality control results have either met or exceeded all CLSI standards.

Karen Campbell
Quality Control Manager