

CERTIFICATE OF ANALYSIS

PRODUCT PP2213SP
TCBS/XLD

LOT NUMBER 1164432

EXPIRY DATE 23.04.2012

Physical Characteristics	Results	Specification
TCBS		
Appearance	Green translucent gel	Green translucent gel
pH	8.65	8.40 - 8.80
XLD		
Appearance	Red translucent gel	Red translucent gel
pH	7.27	7.20 - 7.60

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
TCBS		
O2, 35°C, 24hr		
<i>Vibrio alginolyticus</i> ATCC®17749	2-4mm yellow colonies	
<i>Vibrio parahaemolyticus</i> ATCC®17802	2-3mm green colonies	
<i>Escherichia coli</i> ATCC®25922	Complete inhibition	
XLD		
O2, 35°C, 24hr		
<i>Salmonella typhimurium</i> ATCC®14028	1-2mm pink cols with black centres	
<i>Shigella flexneri</i> ATCC®12022	1-2mm pink colonies	
<i>Escherichia coli</i> ATCC®25922	Complete inhibition	
<i>Enterococcus faecalis</i> ATCC®29212	Complete inhibition	

Released By



Gan May Ping

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com