

XTerminator™ Solution (800mL)

Product No. 4376494
 Lot No. 1205036
 Date of Manufacture 15 Jun 2012
 Expiration Date 25 Nov 2012

TEST	SPECIFICATION	RESULT
------	---------------	--------

PHYSICAL INSPECTION

Label Information	Label Correct and Intact	PASS
Container Integrity	Caps Sealed	PASS

PRODUCT APPEARANCE

Appearance	Light Brown, Opaque Suspension	PASS
------------	--------------------------------	------

USE TEST

A DNA sequencing reaction prepared with BigDye® Terminator v.3.1 chemistry and 100ng of pGEM template was purified with XTerminator™ Solution/SAM™ Solution. The purified reaction was analyzed using an AB 3130xl Genetic Analyzer with a 50 cm capillary and POP-7™ polymer.

% of capillaries exhibiting zero blobs between 60 and 100 bp	>=85%	PASS
--	-------	------

For Research Use Only. Not For Use in Diagnostic Procedures.

**ISO9001:2000
REGISTERED**

Foster City, CA
 Pleasanton, CA
 Bedford, MA
 Warrington, UK
 Singapore



Quality Assurance
Representative

850 Lincoln Centre Drive Foster City, CA T: 650.570.6667 - F: 650.572.2743

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com