



## CERTIFICATE OF ANALYSIS

Page 1 of 1

**PRODUCT** TM0543  
POTASSIUM TELLURITE 1% 100ML

**LOT NUMBER** 1253732

**EXPIRY DATE** 2013.10.05

**RELEASE DATE** 2013.04.07

---

Physical Testing	Results	Specification
------------------	---------	---------------

---

*Tested in accordance with Oxoid Method QC13 unless otherwise stated*

Appearance (Colour)	Pass	Colourless
Clarity	Pass	Translucent
Sterility @ 30°C for 2 days	Pass	Within Limits

*This test is performed in accordance Oxoid method QC21, NATA FADs, ASM Guidelines: Normal inspection AQL 2.5,*  
General inspection level=1

---

Microbiological Testing	Result	Specification	Method
Coryn. diphtheriae gravis ATCC® 19409	Pass	Black colonies	QC27
Coryn. diphtheriae intermedius ATCC®1477	Pass	Black colonies	QC27
Staphylococcus epidermidis ATCC® 12228	Pass	Inhibited	QC27

---

The results reported were obtained at the time of release. Products microbiologically tested are inoculated with bacterial suspension containing either greater or equal to 10,000 cfu for non-target organisms or less than or equal to 100 cfu for target organisms.

Results are recorded semi-quantitatively in comparison to an appropriate control medium. A pass for a target organism is 80-100% recovery (4-5+ growth). A pass for a non-target organism is less than 20% recovery (inhibited, less than 1+ growth).

This product is stable to the stated expiry date when stored in accordance with instructions as per product label.

Released by: Rina Pattullo

.....  
**Approved Signatory**

Accredited for compliance with ISO/IEC 17025.  
This report may not be reproduced except in full.



**MANUFACTURED AND TESTED BY: Oxoid Australia Pty Limited**

20 Dalgleish Street, Thebarton, South Australia 5031

Tel: Toll Free 1 800 33 11 63, Fax : Toll free 1 800 007054

Email: [oxoid.au@thermofisher.com](mailto:oxoid.au@thermofisher.com) Website: [www.oxoid.com.au](http://www.oxoid.com.au)