

CERTIFICATE OF ANALYSIS

PRODUCT PP1745P90
YM 11 AGAR (25ML FILL)

LOT NUMBER 1397044

EXPIRY DATE 31.10.2013

Physical Characteristics	Results	Specification
Appearance	Clear to trace hazy, yellowish brown gel	Clear to trace hazy, yellowish brown gel
pH	7.20	6.80 - 7.20

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
O2, 25°C, 48-52hr		
<i>Aspergillus niger</i> ATCC®16404	Growth, light pale colonies	
<i>Saccharomyces cerevisiae</i> ATCC®9763	Growth, 1.5mm blue grey colonies	
<i>Bacillus subtilis</i> ATCC®6633	Complete inhibition	
<i>Escherichia coli</i> ATCC®25922	Complete inhibition	

Released By



Nurliyana binti Nazri

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com