

Certificate of Analysis



3730 Buffer(10X) with EDTA 500mL

Product No. **4335613**
Lot No. **1505250**
Date of Manufacture **31MAR2015**

TEST	SPECIFICATION	RESULT
VISUAL INSPECTION		
Clear, particulate free, viscous solution		Pass
PH TEST		
Measured at 23.0C ± 0.5C		
pH Value	7.90 - 8.14	7.980
CONDUCTIVITY		
Measured at 23.0C ± 0.1C		
Conductivity of a 1:10 dilution	1500 - 1800 µS/cm	1704 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Bedford, MA, USA

**Quality Assurance
Issued 05JUN15**

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com