



## Certificate of Quality

**Product Name:** ACT II Transport PK/10  
**Lot Number:** 349863

**Product Number:** R12402  
**Expiration Date:** 2014-05-15  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Results</b>
Streptococcus pyogenes	19615	Good Recovery
Bacteroides fragilis	25285	Good Recovery

### Media Physical Characteristics:

Appearance: White, Clear Agar Butt  
pH @ 20-25°C: 7.2 +/- 0.2

### Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Weddle".

Quality Assurance Supervisor