



## Certificate of Quality

**Product Name:** Cary-Blair w/Ind(15ml Vial) PK/120  
**Lot Number:** 439885

**Product Number:** R21925  
**Expiration Date:** 2015-06-16  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Campylobacter jejuni	33291	Good Recovery
Salmonella enterica	14028	Good Recovery
Shigella flexneri	12022	Good Recovery

**Physical Characteristics:**

Appearance: Red, Clear Viscous Liquid  
pH: 8.0 +/- .5

**Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

A handwritten signature in black ink that reads "Susie Waddell".

Quality Assurance Supervisor