

Certificate of Quality

Product Name: PEA w/5% Sheep Blood 100/PK
Lot Number: 731978

Product Number: R01662
Expiration Date: 2009-02-11
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Streptococcus pyogenes	19615	Growth
Enterococcus faecalis	29212	Good Growth
Staphylococcus aureus	25923	Good Growth
Escherichia coli	25922	Marked to Complete Inhibition
Proteus mirabilis	12453	Marked to Complete Inhibition

Physical Characteristics:

Appearance: Red, Opaque Agar
pH @ 20-25°C: 7.3 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor