

CERTIFICATE OF ANALYSIS

PRODUCT TM1246T3
PEPTONE WATER (3ML)

LOT NUMBER 997689


EXPIRY DATE 11.04.2011

Physical Characteristics	Results	Specification
Appearance	Light amber, clear solution	Light amber, clear solution
pH	7.10	7.00 - 7.40

Sterility: Samples of this batch have been incubated at 35°C and room temperature to ensure product meets acceptable levels of sterility.

Microbiological Characteristics	Results	Specification Quantitative Tests
O2, 35°C, 24hr		
<i>Escherichia coli</i> ATCC®25922	Positive for Indole Test	
<i>Klebsiella pneumoniae</i> ATCC®13883	Negative for Indole Test	

Released By



Aishah Dahlan

MANUFACTURED AND TESTED BY: Thermo Scientific Microbiology Sdn Bhd (409466A)

(Formerly known as Biomedica Laboratories Sdn Bhd)

6 Jalan TTC 6 Cheng Technology Park, Cheng, Malacca 75250 Malaysia

Tel: (60)6 334-0975, Fax : (60)6 334-1927 QC Email: oxid.myqcbml@thermofisher.com