

Certificate of Analysis

MethylCode™ Bisulfite Conversion Kit – 50 reactions

Product No. MECOV50
Lot No. IVGN183650
Date of Manufacture 17Jul2015

Package, and label check:

All components of the kit must be present, labeled correctly and contain the appropriate amount of product.

Results: Meets specifications.

Bisulfite Conversion Analysis:

One pg of pUC19 DNA with 600 ng of carrier DNA is bisulfite converted. PCR analysis is then performed using cytosine to thymine specific primers giving a 466 bp amplicon which is visualized on a 1% agarose gel. A no template control shows no amplification as visualized on a 1% agarose gel.

Results: Meets specifications.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Life Technologies
5791 Van Allen Way
Carlsbad, CA, USA 92008
www.lifetechnologies.com
For inquiries, contact us at cofarequests@lifetech.com



Nichole Reaksecker
Quality Manager
Issued on 30 Jul 2015