

Certificate of Analysis



POP 7 Performance Optimized Polymer, 28mL


Product No. 4332241
Lot No. 1701565
Date of Manufacture 13MAR2017
Expiration Date 30OCT2017

TEST	SPECIFICATION	RESULT
VISUAL INSPECTION		
Clear, particulate free, viscous solution		
Appearance		Pass
FUNCTIONAL TESTING		
Each bulk lot must pass all functional performance specifications including KB Q20 quality score.		
Read Length		Pass
FLUORESCENCE TEST		
The ratio of the background fluorescence intensity of the polymer to that of water is measured for each bulk lot to ensure conformance to manufacturing specifications.		
Fluorescence	≤ 2.5	1.4115

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Bedford, MA, USA

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com


Quality Assurance
Issued 15MAR17