

Certificate of Analysis

BenchMark™ His-tagged Protein Standard

Product No. LC5606
Lot No. 1929085
Date of Manufacture 23Jan2018
Expiration Date 30Sep2018

Gel Testing

Ten major sharp bands must be present on a 4–20% Tris–Glycine gel and a 10% Bis–Tris gel with MES buffer. Gel test performance must be comparable to a previously released lot.

Intensity

Band intensity when stained with Invision™ His-tag In-gel Stain must not differ more than 20% from a previously released lot.

Concentration

Ten bands must be visible on the gels when 2 µl of the standard was loaded and stained with the Invision™ His-tag In-gel Stain.

Migration

Bands migrate comparable to a previously released lot.

Contamination


All contaminant bands must be less than 20% of the intensity of the major bands.

Overall Testing

Pass

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Anita Targosz
QA Associate Director
Issued on 30 Jan 2018