



Certificate of Analysis

CloneMiner™ II cDNA Library Construction Kit

Product No. A11180-01

Lot No. 2015911

Date of Manufacture 19Jan2019

Titer Determination:

Specification: Total cfu from Test cDNA BP reaction must be $\geq 1.0 \times 10^7$ cfu.

Result: Meets specifications.

PCR Full Length Analysis:

Specification: $\geq 87\%$ of 24 colonies must have an insert > 500 bp.

Result: Meets specifications.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008

www.thermofisher.com

For inquiries, contact us at cofarequests@thermofisher.com

Anita Targosz
QA Associate Director
Issued on 19 Jan 2019