

Certificate of Analysis

10 mM dNTP Mix
PCR Grade

Product No. 18427-013
Lot No. 2036718
Date of Manufacture 09Nov2018
Expiration Date 30Nov2020

PCR Functional Assay

Specification: PCR reactions containing lambda, ΦX174 RF, and pUC19 DNA as target must yield product of the appropriate size as visualized by gel electrophoresis.

Result: Meets specification.


DNA Contamination Assay

Specification: Amplification reactions lacking target template DNA must show no evidence of product amplified from universal primers when visualized by gel electrophoresis.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Anita Targosz
QA Associate Director
Issued on 30 Nov 2018