

Certificate of Analysis

FreeStyle™ MAX Reagent

Product No. 16447100
Lot No. 2185360
Date of Manufacture 06Mar2020
Expiration Date 31Oct2021

Caution

The safety and efficacy of this product in diagnostic or other clinical uses has not been established.

Sterility

Material is processed through a 0.2 µm filter.

Functional Assay


CHO cells are transfected with a *lacZ* reporter plasmid. The ratio of gene expression between the test lot and a reference lot must be 1.0 ± 0.15 . The peak activities of the test and reference lots must be within ± 0.05 µl.

Results

Meets all specifications

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Anita Targosz
QA Associate Director
Issued on 06 Mar 2020