

## Certificate of Analysis

### ChargeSwitch®-Pro Plasmid Miniprep Kit

Product No. CS30050  
Lot No. 2198242  
Date of Manufacture 21Jun2020

---

#### Yield

Specification: The average (n=5) DNA yield must be  $\geq 2$   $\mu$ g per ml of overnight culture, measured by PicoGreen analysis.

Result: Meets specification.

#### Purity

Specification: Agarose gel electrophoresis of the purified DNA must produce a band of the expected size, as compared to control DNA. Genomic DNA (high molecular weight, >20kbp) and RNA (low molecular weight, <1kbp) should not be present.

Result: Meets specification.

---

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific  
Life Sciences Solutions  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)



---

Shannon Orr  
Sr. Manager, Quality  
Issued on 15 Jul 2020