

Certificate of analysis

ProcartaPlex

Cat. number: **EPX010-12279-901**
96 tests/1 analyte

Name: **ProcartaPlex Human Cystatin C Simplex**

Lot number: **222927-001**

Expiry date: **2021-04**

Components

Components	Quantity	Lot	Store at
BD12279EX Cystatin C Mag Beads	1 x 120µl (50x)	159439000	2-8°C
BK12279EX Det. Ab. Cystatin C	1 x 70µl (50x)	159440000	2-8°C
S14808EX Standard Mix N2	2 each	160746808	2-8°C
UAB10x/11 Universal Assay Buffer 10x	1 x 10ml	186867000	2-8°C

Cystatin C Mag Beads Lot#159439000

Target Name	Bead Number	Std1 Concentration pg/ml	Standard
Cystatin C	34	107800	Standard Mix N2

Standard Mix N2 Lot#160746808

Provided below is a table of Standard 1 (Std1) value for each analyte in each tube when prepared according to the "Preparing Standard" procedure of the Manual.

Analyte	Std1 Concentration (pg/ml)	ULOQ / LLOQ (pg/ml) Determined in cell culture medium
Elafin	7800	1950 / 1,90
REG3a	7950	7950 / 1,94
Lactoferrin	214400	53600 / 52
CD14	129700	129700 / 32
MBL	302200	302200 / 74
CD44	137100	137100 / 33
Cystatin C	107800	107800 / 26
NRP-1	361400	361400 / 88

Analytical information:

This product has been tested by Quality Control and passed internal specifications.

Quality control:

A handwritten signature in black ink, appearing to read "J. J. J. J.", positioned to the right of the "Quality control:" text.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).
For inquiries, contact us at "thermofisher.com/askaquestion"