

Certificate of Analysis

No-Stain™ Protein Labeling Reagent

Product No. A44449
Lot No. 2266948
Date of Manufacture 23Oct2020
Expiration Date 26Aug2021

Linearity

Specification: Following quantification of a dilution series of protein in a gel transferred to a membrane and labeled using the No-Stain™ Protein Labeling Reagent, analysis shall yield an R^2 value ≥ 0.90 for linear fit of ≥ 4 data points.

Result: Meets specification.

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).

Thermo Fisher Scientific
Life Sciences Solutions
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com



Shannon Orr
Sr. Manager, Quality
Issued on 23 Oct 2020