

pGEM3Zf (+) DNA

Product No. **360451**
Lot No. **1710171**
Date of Manufacture **20OCT2017**

TEST	SPECIFICATION	RESULT
PHYSICAL INSPECTION		
360451_PHYSICAL INSPECTION		Pass
TESTING		
300288_A260/A280 RATIO	>= 1.8	1.9
300288_UV ABSORBANCE AT 260NM	0.70 - 0.90 AU	0.79 AU
300288_PURITY/IDENTITY BY AGAROSE GEL		Pass
300288_A260/A250 RATIO	>= 1.05	1.1

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Austin, TX, USA.

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

**Quality Assurance
Issued 30MAY18**