

SILANE Lysis/Binding/Washing Buffer 1 (viral NA)

Product No. **37031D**
Lot No. **1708037**
Date of Manufacture **18AUG2017**
Expiration Date **01AUG2020**

TEST	SPECIFICATION	RESULT
PH TEST		
pH at 25degC	6.2 - 6.8	6.27
FUNCTIONAL TEST		
Recovery of M13 viral DNA	60 - 136 %	78 %
Bacterial Bioburden		Pass
Fungal Bioburden		Pass

Caution: For manufacturing, processing, or repacking

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Warrington, UK



**Quality Assurance
Issued 17NOV17**

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com