



Certificate of Analysis

StepOnePlus™ Real-Time PCR System Spectral Calibration Kit

Product No. **4371435**
Lot No. **2110361**

TEST	SPECIFICATION	RESULT
Material Test Applied Biosystems uses spectrofluorimetric analysis to test fluorescence emission wavelength maximum in each lot of component bulk material used in the Spectral Calibration Kit.		
SYBR® GREEN Dye	522 ± 5 nm	522 nm Pass
ROX™ Dye	601 ± 5 nm	601 nm Pass
NED™ Dye	570 ± 4 nm	572 nm Pass
JOE™ Dye	548 ± 4 nm	550 nm Pass
VIC® Dye	548 ± 4 nm	550 nm Pass
FAM™ Dye	518 ± 4 nm	517 nm Pass

For Research Use Only. Not for use in diagnostic procedures.



Singapore

Tehai Kai Yen

09 DEC 2021

Quality Assurance



Life Technologies Holdings Pte Ltd, Block 33, Marsiling Industrial Estate Road 3, #07-C5,
Singapore 739256 Tel: (65) 6362 9300

To obtain a Certificate of Analysis on-line go to www.appliedbiosystems.com or email us at cofarequest@appliedbiosystems.com