

Certificate of Analysis

Anode Buffer Container 3500 Series

Product No. **4393927**
Lot No. **1801437**
Date of Manufacture **07FEB2018**
Expiration Date **23OCT2018**

TEST	SPECIFICATION	RESULT
------	---------------	--------

PHYSICAL INSPECTION Material is clear, colorless and free of particulates		
---	--	--

Appearance - Conformance		Pass
--------------------------	--	------

PH TEST Measured at 25.0°C ± 0.3°C		
--	--	--

pH Value	7.6 - 8.2	7.962
----------	-----------	-------

CONDUCTIVITY TEST Measured at 25.0°C ± 0.1°C		
--	--	--

Conductivity	1500 - 2000 µS/cm	1784 µS/cm
--------------	-------------------	------------

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system. Site: Bedford, MA, USA



**Quality Assurance
Issued 12FEB18**

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com