Certificate of Analysis



MicroSEQ[®] ID; ID Sequencing Cleanup Plates

Product No.	4408227
Lot No.	170502
Expiration Date	16MAY2020

TEST	SPECIFICATION	RESULT
Inspection		
Product appearance	White gel slurry, free of particulates	Pass
Functional test 96-well ultra plates were tested for functionality on an ABI 3730 sequencer by sequencing a standard plasmid, pGEM or		
pUC19, and assessing signal intensity, absence of fluorescent contaminants, observed background, and sequencing accuracy.		
Terminator breakthrough	< 5%	Pass
Fluorescent contaminants	None	Pass
Average length of read	800-1100bp	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 9001 and ISO 13485 certified quality management system.

Austin, TX, USA

Cinthia Daona

Quality Assurance Issued 31JAN2019

Life Technologies 5791 Van Allen Way Carlsbad, CA, USA 92008 <u>www.thermofisher.com</u> For inquiries, contact us at <u>cofarequests@lifetech.com</u>.

> FORM-00013E Doc: 4489610C